

RESEARCH ARTICLE

A wearable wrist haptic display for motion tracking and force feedback in the operational space

Marco Laghi^{1,2*} , Manuel G. Catalano¹, Giorgio Grioli^{1,2} and Antonio Bicchi^{1,2}

¹Soft Robotics for Human Cooperation and Rehabilitation, Istituto Italiano di Tecnologia, Genova, Italy

²Centro di Ricerca “Enrico Piaggio,” Università di Pisa, Pisa, Italy

*Corresponding author: Email: marco.laghi89@gmail.com

Received: 24 June 2020; Revised: 12 February 2021; Accepted: 05 April 2021

Keywords: Feedback devices; Design; Control; Mechatronics

Abstract

Force feedback is often beneficial for robotic teleoperation, as it enhances the user’s remote perception. Over the years, many kinesthetic haptic displays (KHDs) have been proposed for this purpose, which have different types of interaction and feedback, depending on their kinematics and their interface with the operator, including, for example, grounded and wearable devices acting either at the joint or operational space (OS) level. Most KHDs in the literature are for the upper limb, with a majority acting at the shoulder/elbow level, and others focusing on hand movements. A minority exists which addresses wrist motions. In this paper, we present the Wearable Delta ($W\Delta$), a proof-of-concept wearable wrist interface with hybrid parallel–serial kinematics acting in the OS, able to render a desired force directly to the hand involving just the forearm–hand subsystem. It has six degrees of freedom (DoFs), three of which are actuated, and is designed to reduce the obstruction of the range of the user’s wrist. Integrated with positions/inertial sensors at the elbow and upper arm, the $W\Delta$ allows the remote control of a full articulated robotic arm. The paper covers the whole designing process, from the concept to the validation, as well as a multisubject experimental campaign that investigates its usability. Finally, it presents a section that, starting from the experimental results, aims to discuss and summarize the $W\Delta$ advantages and limitations and look for possible future improvements and research directions.

Introduction

Force feedback has been repeatedly proved to greatly improve the quality of a teleoperation scenario, for example, in the case of remote interaction (Yip et al., 2011). Indeed, beneficial applications of haptic interfaces to teleoperation can be found in several fields, from telesurgery (Guthart and Salisbury, 2000), to space (Imaida et al., 2004) and underwater exploration (Khatib et al., 2016), as well as disaster response (Klamt et al., 2019a).

Numerous kinesthetic haptic displays (KHDs) have been proposed to deliver a feedback by acting in the domain of forces and displacements. Usually, these assume that the interaction with the remote environment occurs through one main contact interface, that is, the end effector, which can correspond to the palm/back of the hand or a fingertip of the operator and aim at rendering the wrench corresponding to the total of all the forces and torques exchanged through that contact interface. The way a KHD returns this wrench depends mainly on its kinematics and on its coupling with the operator, and the choice of both

© The Author(s), 2021. Published by Cambridge University Press. This is an Open Access article, distributed under the terms of the Creative Commons Attribution-NonCommercial-ShareAlike licence (<http://creativecommons.org/licenses/by-nc-sa/4.0/>), which permits non-commercial re-use, distribution, and reproduction in any medium, provided the same Creative Commons licence is included and the original work is properly cited. The written permission of Cambridge University Press must be obtained for commercial re-use.




Figure 1. Scheme of generalized interaction between operator and kinesthetic haptic display. $f_{c_i}, i = 1, \dots, 4$ are the exchanged forces between the human and the interface at the contact points (assumed to be 4 for simplicity). The force exchanged through the first contact point f_{c_1} is also called f_B , whereas the last (f_{c_4}) is called f_E . The second and third are then collected in the vector $f_C = [f_{c_2}^T f_{c_3}^T]^T$. f_{GH} and f_{GI} denote the contact forces with the ground of the human and the mechanism, respectively.

these elements can strongly influence the outcome of the teleoperation experience. The motivation behind this aspect can be understood with the help of the general user-interface scheme in Figure 1, where a KHD addressing the whole arm workspace is represented.

Traditionally, KHDs were implemented in the form of robotic system with a grounded base (G), to not burden the heavy weight of the system on the operator. With respect to the scheme of Figure 1, G-KHDs always present a ground reaction force f_{GI} . More recently, thanks to the achievements in miniaturization and the use of lightweight materials, engineers started proposing wearable devices (W), that ditch the direct ground connection (f_{GI} absent) to improve portability of the device and mobility of the user. Please note that the following classification based on the device kinematics and its coupling with the user is just one of the possible topologies. Other classifications can be based, for example, on the used actuation, sensing method, and device control. We will not analyze these last, as they are not the focus of this work and have been already analyzed in existing literature, as Rosen and Ferguson (2020) and Pacchierotti et al. (2017).

The majority of existing systems, both grounded and wearable, are designed as exoskeletons that follow the motion of the operator, link by link, with kinematic joints accommodating for the human articulations. We will refer to these designs as “joint-space” (JS). An important characteristic of JS-KHD is that the kinematics of the human and the device are designed to be, as much as possible, equal. Following this assumption, the torque exerted by each joint of the interface is directly mapped on the corresponding joint of the human, through the various intermediate contact forces (see Figure 1), a hypothesis that heavily simplifies the control of the device. However, the hypothesis of matching kinematics is hardly achievable in reality. Indeed, human joints are not easy to reproduce with a mechanical counterpart. Although they are often approximated with pin joints, their motion is never really purely rotational, nor they are exactly the same across different people. Moreover, their alignment with the KHD joints tend to be always imperfect. Also, a rigid and precise attachment between the two is challenging, due to the compliance of human tissues. As a consequence, undesired relative movements and contact forces outside the image space of the expected wrench can occur. The majority of JS systems live with this problem, accepting the generation of such undesired forces, even if they downgrade the interface performance and the quality of the haptic rendering. Finally, a fundamental limitation of JS-KHD is that they cannot render a desired wrench W_d exactly, as they only render its projection onto the space of the joint torques, and transfer it to the operator through the actuatable contact forces.

Plenty of examples of grounded, joint-space displays (GJS-KHD) can be found in literature, involving one or more fingers (Agarwal et al., 2015), wrist (Gupta et al. 2008; Pehlivan et al. 2012; Martinez et al. 2013; Andrikopoulos et al. 2015; Pezent et al. 2017; Basteris et al. 2018; Lee et al. 2018; Pezent et al. 2019), and full-arm (Frisoli et al., 2005; Gupta and O’Malley, 2006; Mallwitz et al., 2015; Barsotti et al.,

2018). Several are also the wearable, joint-space solutions (WJS-KHD), in which f_{GI} is not present, and therefore all the weight of the structure has to be compensated by the user (the interface represented in Figure 1 would start from the contact point corresponding to the force $f_{c1} \equiv f_B$). Nonetheless, wearable devices tend to be lighter than the grounded ones, and, ideally, can be carried and used wherever desired. Similarly to GJS-KHDs, WJS-KHDs can involve different portions of the operator limb. There are some examples of full-arm exoskeletons for teleoperation (Letier et al., 2008; Schiele and Hirzinger, 2011; Klamt et al., 2019b), and some of hand devices as well (Leonardis et al., 2015; Gabardi et al., 2018). On the contrary, there are relatively few wearable JS-KHDs for the wrist and, to the best of the author's knowledge, almost all of them are designed for application in rehabilitation (Hope and McDaid 2017; Lambelet et al. 2017; Buongiorno et al. 2018). Note that in both wearable and grounded JS categories, there are several examples of solutions in which particular attention has been put in designing structures that aims to reduce or even remove the macromisalignment issues, as Schiele and Van Der Helm (2006), Sergi et al. (2012), and Jarrassé and Morel (2011). In Esmaeili et al. (2013), for example, a wrist exoskeleton with customizable joint parameters has been proposed for such purpose.

An alternative to building an exoskeleton device is that of building a system with one main contact interface with the operator, exchanging forces with them not anymore at the level of joint torques, but at the level of the operational space (OS) devices. Differently from JS, OS interfaces bypass the joint alignment issue, acting directly at the end effector, therefore removing all the intermediate contact points and the associated forces f_C . Furthermore, by avoiding to replicate the human kinematics, this class of devices has the potential to obtain a subspace of actuatable contact forces of full rank and to render an exact and complete wrench at the end effector, without intervening at the joint level. Grounded, operational space (GOS) devices are probably the most popular, and some of them are also available in the market. In this case, f_B is absent (and so the only contact force between human and device is f_E in Figure 1). Examples of small-workspace GOSs are the ones commercialized by Force Dimension¹ and 3D Systems,² or other platforms developed in research labs, as Frisoli et al. (2004) and Okamura et al. (2002), as well as the haptic interface of the DaVinci surgical robot from Intuitive.³ An example of larger workspace can be found in Hulin et al. (2011), where seven DoF robotic arms are used as haptic interfaces. An interesting solution can be found in Oblak et al. (2010), which proposes a system that can be configured to work for arm or wrist GOS rehabilitation.

Finally, there is the case of wearable OS devices (WOS-KHD), which try to combine the advantages of GOS-KHD with that of being portable. They present some contact forces at the base f_B , while there are no f_{GI} . Consequently, the interface weight compensation is demanded to the user, as for WJS. Furthermore, when rendering a given f_E , also a f_B is inevitably generated at the base attachment. Nonetheless, as already said, wearable devices offer the undisputed advantage of portability. In the knowledge of the authors, WOS-KHDs are very few, with only few applications for the hand (Frisoli et al., 2007; Fontana et al., 2009; Iqbal et al., 2015; Choi et al., 2016; Sarakoglou et al., 2016), and none for arm and wrist. Table 1 reports the devices just listed, divided in the categories they belong to.

WOS solutions are less explored due to the difficulties in their design (in particular regarding the stability of their anchors to the user). Nonetheless, they are possibly the most interesting among all systems, because they combine several advantages. Indeed, wearability increases portability and facilitates the usage, since WOS-KHDs do not need a particular installation site or support.


Inspired by the advantages just described, this paper explores the possibility to design a new WOS for the wrist to be used in bilateral teleoperation. The idea is to design a light device that embraces the forearm and hand; they can measure the user's wrist movements without obstructing them, and can render a force at the user palm coherent with the interaction sensed at the remote side. The result is a stand-alone WOS wrist KHD named "Wearable Delta" (W Δ), shown in Figure 2. The device is anchored onto the forearm and uses a hybrid parallel–serial kinematic structure to render forces at the user's hand, avoiding the

¹ <http://www.forcedimension.com>.

² <https://www.3dsystems.com/haptics>.

³ <https://www.intuitive.com>.

Table 1. Classification of existing kinesthetic haptic displays and relative examples.

	GROUNDED (G)	WEARABLE (W)
JOINT SPACE (JS)	<div style="display: flex; justify-content: space-around;"> <div style="text-align: center;"> <p>full-arm</p>  </div> <div style="text-align: center;"> <p>wrist</p>  </div> <div style="text-align: center;"> <p>finger</p>  </div> </div> <p>Barsotti et al. (2018) Gupta et al. (2008) Agarwal et al. (2015) Frisoli et al. (2005); Mallwitz et al. (2015); Gupta and O'Malley (2006); Martinez et al. (2013); Andrikopoulos et al. (2015); Lee et al. (2018); Pehlivan et al. (2012); Pezent et al. (2017,0); Basteris et al. (2018).</p>	<div style="display: flex; justify-content: space-around;"> <div style="text-align: center;"> <p>full-arm</p>  </div> <div style="text-align: center;"> <p>wrist</p>  </div> <div style="text-align: center;"> <p>finger</p>  </div> </div> <p>Letier et al. (2008) Lambelet et al. (2017) Gabardi et al. (2018) Schiele and Hirzinger (2011); Leonardis et al. (2015); Hope and McDaid (2017); Buongiorno et al. (2018); Klamt et al. (2019b).</p>
OPER. SPACE (OS)	<div style="display: flex; justify-content: space-around;"> <div style="text-align: center;">  <p>Omega 7 by Force Dimension</p> </div> <div style="text-align: center;">  <p>Hulin et al. (2011)</p> </div> <div style="text-align: center;">  <p>DaVinci by Intuitive</p> </div> </div> <p>Frisoli et al. (2004); Okamura et al. (2002); Oblak et al. (2010).</p>	<div style="text-align: center;"> <p>only finger</p> <div style="display: flex; justify-content: space-around;">   </div> <p>Sarakoglou et al. (2016) Choi et al. (2016) Fontana et al. (2009); Iqbal et al. (2015); Frisoli et al. (2007).</p> <p style="color: red;">for the wrist</p>  <p>The proposed wrist KHD Wearable Delta (WΔ- Fig. 2)</p> </div>




Figure 2. Picture of the $W\Delta$ worn, complete with sensorized elbow brace and Myo armband for the arm trajectory tracking: (a) front and (b) back views.

problem of wrist axis joint misalignment. It also integrates a sensorized elbow brace and an inertial measurement unit (IMU) that, combined with the position sensors of the mechanism, can be used to calculate the user's arm pose and their hand trajectory, which can in turn be used as input to the remote platform.

The paper is structured as follows: Section "Concept" introduced the concept of the $W\Delta$ and briefly summarizes its forward kinematics, exhaustively analyzed in Appendix A. The design specifications and implementation follow in Section "Design." Section "Characterization" reports the mechanical characterization of the $W\Delta$, and Section "Psychophysical Assessment" describes a performed psychometric


assessment. Section “Teleoperation Framework” quickly summarizes the implemented bilateral teleoperation scheme (its complete analysis is reported in Appendix B), divided in the delivery of haptic feedback and the control of a remote robotic arm. Section “Experimental Validation” reports two example tasks to show general functionality, whereas Section “Usability Assessment” reports a preliminary usability assessment of the device and Section “Discussion, Limitations, and Future Developments” discusses the results and possible future improvements. Finally, in Section “Conclusion,” we conclude this manuscript with a critical summary of results and a description of future work.

Concept


The idea proposed in this work is to design a WOS-KHD attached to the forearm that acts directly on the palm for bilateral teleoperation applications. The wearable design is chosen to improve its portability, and the involvement of the final portion of the user arm only limits the device total weight and size. The choice of the OS option, instead, is motivated to improve the rendering precision of the desired wrench directly to its application point, in order to augment the haptic sensation and to make it as close as possible to the remote interaction. To avoid any possible kinematic mismatch issues, the device has six DoFs, since a frame fixed in the palm center performs both rotations and translations with respect to the forearm. The choice of the kinematic structure considers also user’s comfort. In particular, the chosen kinematics influence the weight distribution on the arm, and consequently the effort required for its compensation. A structure with a good weight distribution is preferable, for example, minimizing the distance of its center of mass with respect to the forearm longitudinal axis. Symmetric parallel kinematic chains can help matching this requirement. Some of the wrist devices listed in Section “Introduction” already implement parallel structures, from which take inspiration, as Andrikopoulos et al. (2015) and Lee et al. (2018). The most interesting, and the one from which our device takes partial inspiration, is the RiceWrist. Presented in Gupta et al. (2008) and Gupta and O’Malley (2006), it is a GJS rehabilitation device with four DoFs which uses a specific parallel mechanism, introduced in Lee and Shah (1988), with two rotational DOFs and one translational DoF. The design specifications of the RiceWrist, which is intended for rehabilitation, are such that the DoF arrangement and kinematics are thought to precisely guide the patient movements during the rehabilitation process. Nonetheless, a solution that foresees the explicit application of forces in function of trajectory errors and assistance-needed parameters has been proposed in Pehlivan et al. (2014). This indicates that haptics and rehabilitation devices can converge to similar control solutions, and mainly differ in the intended application and consequent design specifications. On this aspect, our wrist KHD is not conceived and designed to perform movement guidance, but rather to exert desired forces (coherent with a remote interaction) while allowing unconstrained user wrist motion and offering complete wearability.

The solution we propose is a structure resulting from the combination of both parallel and serial chains. It consists in the serial assembly of two quite common structures (see Figure 3a): a Delta robot structure (Clavel, 1988) and a three-DoF gimbal. As shown in Figure 3a, the base of the Delta is conceived as a ring to fit the upper part of the forearm, with the actuators distributed around it. Ideally, when the hand is aligned with the forearm, the moving plate center coincides with the hand palm center. The Delta structure has three DoFs, all linear. Consequently, without additional joints, the palm’s rotations would be blocked, and so any wrist movement. The addition of a gimbal structure on the Delta’s plate (brown in Figure 3a) provides for the three missing rotational DoFs. The moving plate is reduced to just a ring used to connect each of the three legs universal joints to the external arc of the gimbal. The gimbal, then, uses the ring as a base, with its axes passing through the palm’s center. Its last link is fixed on the user’s backhand. Providing actuated joints also for the gimbal would mean adding more weight at the user’s arm tip, and a less compact solution, due to the space required by the actuators. Therefore, the idea is to leave the gimbal passive and to rely only on the Delta actuation, which will be controlled in torques for the haptic feedback. It follows that the device is able to provide linear forces only at the backhand anchor point. Please note that, with this kinematics, the device will automatically adapt to the user’s arm lengths. Indeed, while the gimbal enables the wrist rotation, the Delta structure automatically adapts to the user arm (especially along the forearm axis), as it opposes resistance to hand movements only when a force actuation is commanded.

- Delta structure
- Active joints
- Elbow brace flange
- Gimbal structure
- Passive joints
- Elbow encoder support
- Electronic boards



(a)



(b) From top to bottom: gimbal, leg, and base.

	Base		Leg			Gimbal				
Dimension	$r_{b\Delta}$	$r_{b\Delta,int}$	L_{Δ}	l_{Δ}	l_{fb}	$z_{p,off}$	$r_{p\Delta}$	$r_{g,int}$	$r_{g,mid}$	$r_{g,ext}$
Value [m]	0.0975	0.07	0.16	0.16	0.045	0.025	0.0975	0.075	0.09	0.1105

(c)

Figure 3. Technical drawings and information of the W_{Δ} : (a) real device aspect with its different functional components highlighted; (b) tech drawings of the three main parts with highlighted lengths defined in the device; and (c) table summarizing the chosen values of the different dimensions interested in the design.

Furthermore, since the W_{Δ} totally envelopes the user’s forearm and hand, its kinematics can be used to reconstruct the pose and velocities of the user’s arm. The study of the device kinematics and its integration with the humans is explained in detail in Appendix A. For the sake of brevity, we report here the final results. The forward kinematics of the user arm which exploits the W_{Δ} kinematics can be written as follows:

$$H_{pm}^0(q_H) = H_{sh}^0(q_{sh})H_{el}^{sh}(q_{el})H_{b_{\Delta}}^{el}H_{pm}^{b_{\Delta}}(q_{W_{\Delta}}), \tag{1}$$

with H_{pm}^0 being the homogeneous transformation matrix of the palm frame Ψ_{pm} with respect to world fixed frame Ψ_0 .

Similarly, the system Jacobian, which allows to express the user palm twist as function of the arm and device joint velocities, is

$$T_{pm}^{pm,0} = J_{\overline{H}}^{pm} (6 \times 10) \dot{q}_{\overline{H}} = \begin{bmatrix} J_{H_{el}}^{pm} (6 \times 4) & J_{W\Delta}^{pm} (6 \times 6) \end{bmatrix} \begin{bmatrix} \dot{q}_{sh} \\ \dot{q}_{el} \\ \dot{q}_{W\Delta} \end{bmatrix}. \quad (2)$$

For a detailed explanation of these formulas, please refer to Appendix A. In the next section, the design of the $W\Delta$ is reported.

Design

This section presents the hardware design of the $W\Delta$. First, the description of the mechanical requirements and their accomplishment will be explained, followed by the description of the electronics, sensors, and actuators.

Mechanical Design

The technical drawing of the realized device are depicted in Figure 3. Since the $W\Delta$ has to be worn and attached to the forearm, the base of the Delta structure is designed as a ring (see Figure 3a,b). Referring to Figure 3a, the Delta base holds the actuators (cases in red) and the electronics (cases in dark gray). The forearm anchor is created linking the Delta base to an elbow brace (see Figures 2 and 3a for the elbow brace flange, depicted in dark green). The reason of this is to overcome to the stretch and slip of the skin. Embracing both arm and forearm and providing a rigid plate, the brace helps the $W\Delta$ to maintain a stable position despite the arm movements. The elbow encoder support (light green in Figure 3a) is attached to the brace flange. All these components are 3D printed in acrylonitrile butadiene styrene (ABS).

The kinematics of each leg of the Delta is R–U–U. It has been realized with the rotational joint centered at the output shaft of each actuators, while an equivalent solution to the two universal joint has been created combining two rotational joints parallel to the actuator one and a four-bar mechanism with its joint axes normal to the previous (Figure 3). All the parts of the legs, in yellow in Figure 3a, are realized in stainless steel, whereas the joints (light blue) are in aluminum.

The moving plate of the Delta, that also works as base for the gimbal, is designed as a ring as well. It is supported by the last joints of the Delta legs, and in turn it houses the first joints of the gimbal, which is simply realized as a layering of rotating arcs. The most internal arc is designed to match with the hand support made for the commercial GoPro⁴ action camera called *The Strap*,⁵ to which the user's backhand is anchored. In this way, the user does not have to grasp any handle, as the interface is fixed at the back of its hand through straps (see the video linked in the footnote to easily understand *the Strap* functioning). The moving plate and the gimbal arcs (brown in Figure 3a) are 3D printed in ABS, while the gimbal passive joints (light blue) are realized in aluminum and house the gimbal encoders.

Actuators and Electronics

The motors used to actuate the three base joints of the Delta structures are Maxon Motor DCX 22 S + GPX22 12 Volt (83:1). This combination allows a maximum output continue torque of 1.2118 Nm at 1.65 A, and a peak torque of 8.964 Nm at 11.8 A. The chosen motor and gearbox have the efficiency of 85.2 and 74%, respectively, which is traduced in a very good backdrivability. This aspect is very important in a teleoperation scenarios, as the device should be as transparent as possible. To measure the rotation of the different joint axes, we use Austrian Microsystems 5054 position sensors, with a 16-bit resolution, corresponding to a resolution of $\sim 0.011^\circ$ ($\approx 2 \times 10^{-4}$ rad). A total of seven encoders is present: three placed coaxially to the output motor shafts, three in the gimbals, and one

⁴ <https://gopro.com>.

⁵ https://www.youtube.com/watch?v=mdigZJL7hMQab_channel=GoProTips.

aligned with the rotational axis of the elbow brace (see Figures 2 and 3a). As a consequence, the measurement resolution of the gimbal axis position (wrist rotation) is the same of the used encoders. The Delta plate position measurement resolution, instead, is not constant, as it depends on the Delta Jacobian (see (17)), and therefore on its dimensioning (explained hereafter) as well as its instantaneous joint position. Moreover, it also varies depending on the movement direction. Just to give an idea, considering the link length resulting from the following dimensioning, a standard user 1.8-m tall, and a straight wrist position ($R_{pm}^{b\Delta} = I$ and $p_{p\Delta}^{b\Delta} = [0, 0, 0.2508]$ m), we have a movement resolution of ~ 0.3 mm along x , ~ 0.12 mm along y , and ~ 0.05 mm along z . Given the upper-arm length l_u (Figure 17b), the elbow encoder is used to calculate the matrix H_{el}^{sh} (see (21) in Section “Human Arm Kinematics” in Appendix A). The electronic boards used to read the encoders and control the motors are the ones presented in Della Santina et al. (2017).

At the upper arm, a Myo armband is worn (see Figure 2). In the Myo armband, an IMU is used to extrapolate the rotation matrix of the homogeneous transformation matrix H_{sh}^0 (see Section “Human Arm Kinematics” in Appendix A). Moreover, the user holds a small trigger, embedding an eighth encoder, which can be used to control the end effector of the slave, for example, opening and closing a gripper.

Device Dimensioning and Anthropomorphic Matching

The first general requirement for the $W\Delta$ is to be able to fit different arm sizes. From Winter (2009), the forearm length (from elbow to wrist) of an adult is the 14.5% of its height, while the hand length (from wrist to fingertips) is the 10.8%. These estimates are valid for both males and females. Then, covering users with height ranging $h = [1.6, 2]$ m, and adjusting the rounding up so to match also outliers, the forearm length ranges can be set to $l_f = [0.23, 0.32]$ m. The range distance between the wrist’s and the palm’s center (calculated as a third of wrist–fingertips length) can be set to $l_h = [0.06, 0.08]$ m. The workspace the device has to cover can be defined by combining the lengths l_f and l_h and the range of motion shown in Table 2. It has to be considered also that the Delta base will be anchored on the forearm at a distance of ~ 0.05 m from the elbow. Finally, the desired workspace can be rounded up joining all the upper half sphere centered in $z = l_f - 0.05$ m with a radius $\rho = l_h$, with $l_f = 0.145h$ and $l_h = \frac{0.108}{3}h$.

All the lengths chosen in the design analysis that will follow are summarized in Figure 3c. The Delta structure parameters are basically four (see Figure 17a): (i) the radius of the base $r_{b\Delta}$; (ii) the radius of the moving plate $r_{p\Delta}$; (iii) the length L_Δ of the driver leg; and (iv) the length l_Δ of the driven leg.

Table 2. Range of motion (ROM) of human (a) forearm supination/pronation, (b) wrist abduction/adduction, and (c) wrist flexion/extension and length ranges of (d) forearm and (e) hand (wrist–palm’s center) human segments

	(a) sup/pro	(b) abd/add	(c) flex/ext
ROM	85°/85°	19°/45°	90°/90°
	(d) Forearm l_f		(e) Hand l_h
Length range	[0.23, 0.32] m		[0.06, 0.08] m

Table 3. Device torques and forces summary. Columns divided in: (a) maximum continuous and (b) peak values (both from data sheet), and (c) used ones (through software saturation)

	(a) Max. continuous	(b) Peak	(c) Max. used
Motors’ current (A)	1.65	11.8	1.5
Motors’ output torques (Nm)	1.2118	8.964	1.101
Force at $\Psi_{p\Delta}^a$ (N)	5.5	39.3	5.5
Torque at wrist (Nm; with $l_h = 0.07$ m)	0.385	2.751	0.35

^aThis is the maximum force assured in all the interested workspace given the current saturation at 1.5 A, as specified in Section “Device Dimensioning and Anthropomorphic Matching.” This means that, depending on the device Jacobian (and thus on its joint position), it is also possible to reach greater values.

The two radii are chosen in order to fit all the arm sizes. In particular, $r_{b_{\Delta}}$ is influenced by the internal radius of the base circle $r_{b_{\Delta,int}}$ (see the third drawing of Figure 3b), that should be large enough to fit any forearm. It has been set $r_{b_{\Delta,int}} = 0.07$ m and, as consequence, $r_{b_{\Delta}} = 0.0975$ m.

On the other side, the gimbal has to be large enough to not interfere with the user’s wrist in all its movement. Considering this, the various radii of the gimbal have been set as (see top drawing of Figure 3b): $r_{g,ext} = 0.1105$ m, $r_{g,mid} = 0.09$ m, and $r_{g,int} = 0.075$ m. In this way, the hand should never touch the structure.

Desiring a compact and easily foldable solution, we would like to have $r_{p_{\Delta}} + l_{\Delta} = r_{b_{\Delta}} + L_{\Delta}$, and in particular $r_{p_{\Delta}} = r_{b_{\Delta}}$ and $l_{\Delta} = L_{\Delta}$. The first condition ($r_{p_{\Delta}} = r_{b_{\Delta}}$) is achieved by shifting the plane in which the plate joint legs lie below the gimbal structure (an area never occupied by the user’s limb), with an offset of $z_{p,off} = 0.025$ m (see the gimbal drawing of Figure 3b).

Once $r_{b_{\Delta}}$ and $r_{p_{\Delta}}$ have been defined, we would like to define the lengths L_{Δ} and l_{Δ} that (i) cover the desired workspace (the one spanned by the user palm center) in the one of W_{Δ} and (ii) assure the rendering of a continuous force of at least 5 N in any direction in any point of the desired workspace, considering a maximum continue current of 1.5 A (slightly lower than the one provided by data sheet in order to be sure to prevent motors’ damage). The possible combinations of l_{Δ} and L_{Δ} are explored through a recursive algorithm, which uses the inverse kinematic algorithm proposed in Williams (2016), considers also the offset $z_{p,off}$, and varies the lengths L_{Δ} and l_{Δ} with steps of 0.01 m. Finally, we choose a combination of L_{Δ} and l_{Δ} that respect the second condition $l_{\Delta} = L_{\Delta}$, minimizing these two lengths: $L_{\Delta} = l_{\Delta} = 0.16$ m. This choice reduces the envelope of the device as much as possible and, together with $r_{p_{\Delta}} = r_{b_{\Delta}}$, also the complete fold and compactness of the interface when not used. The width of the four-bar is set as $l_{fb} = 4.5$ cm. Table 3c summarizes all the values chosen for the different length of the W_{Δ} , whereas Table 4 summarizes the currents’ and motors’ torques from data sheet, as well as the corresponding assured force at $\Psi_{p_{\Delta}}$ and the torque at the user wrist in the nominal case $l_h = 0.7$ m. Last column, instead, show the various data considering as maximum current the used value of 1.5 A, as explained just above.

The W_{Δ} itself has a weight of 1.775 Kg, which grows to 2.224 Kg when considering the elbow brace (0.356 Kg) and the Myo armband (93 g). Figure 2 shows the actual final result mounted on a user’s arm.

Characterization

A characterization of the device was performed, in order to match the actual force applied at the end effector with the desired one. The setup used for the W_{Δ} characterization is depicted in Figure 4. The passive gimbal is unmounted from the W_{Δ} , consequently leaving only the Delta part, and replaced with a structure that functions as a flange between the Delta and the tool side of an ATI mini45-e F/T sensor.⁶ The sensor is fixed to the end effector of a Panda robotic arm (Franka Emika⁷). The orientation of the Panda end effector is fixed so to be normal to the Delta base for the whole characterization. Its position is instead regulated so to match the desired position of the Delta moving plate center.

The characterization is performed as follows: 38 points lying on the edge and inside the expected users’ workspace are chosen. In each point, 144 directions on a unitary sphere are defined. All these directions define the versors $\hat{f}_d \in \mathbb{R}^3$ of the desired forces. For each direction, a force is applied with intensity $\|f_d\|$ going from 0 to 5 N and then back to 0 again, with steps of 1 N applied for 1.5 s. The resulting force is then

$$f_d = \hat{f}_d \|f_d\|. \tag{3}$$

Through the Delta Jacobian $J_{\Delta}(\theta_{\Delta})$ of (17), the desired torques $\tau_{\Delta,d}$ are calculated as

⁶ https://www.ati-ia.com/products/ft/ft_models.aspx?id=Mini45.

⁷ <https://www.franka.de>.




Figure 4. Setup used for the characterization: the Delta has the base anchored to the table, and the plate attached to the Franka robot end effector through the ATI mini45-e F/T sensor. The Franka end-effector rotation is blocked so to be always look perpendicular to the table plane, while it is moved to pose the Delta plate in the interested position.

$$\tau_{\Delta,d} = J_{\Delta}(\theta_{\Delta})^T f_d, \tag{4}$$

and the commanded currents to the motors I_c are calculated as

$$I_{\Delta,c} = C_{T_m,ds} \tau_{\Delta,d}, \tag{5}$$

being $C_{T_m,ds}$ the current–torque actuator constant provided by the data sheet. For all the 38 chosen points, the commanded currents $I_{\Delta,c}$ to the motors, the motors’ positions θ_{Δ} , and the force measured by the F/T sensor f_m are registered and averaged in the time window of the application of all the desired forces. The average measured force \bar{f}_m is then converted in measured torques $\tau_{\Delta,m}$ through the Delta Jacobian:

$$\tau_{\Delta,m} = J_{\Delta}(\theta_{\Delta})^T \bar{f}_m. \tag{6}$$

Figure 5a–c shows the 3D plots of $\theta_{\Delta} - \tau_{\Delta} - I_{\Delta}$. Projected on the $\tau_{\Delta} - I_{\Delta}$ plane, the three characteristics can be well approximated by a straight line. The resulting $\tau_{\Delta} - I_{\Delta}$ relations for each motor are then approximated as

$$I_{\Delta_1} = 1.308\tau_{\Delta_1} + 0.009552; \quad I_{\Delta_2} = 1.462\tau_{\Delta_2} - 0.002873; \quad I_{\Delta_3} = 1.436\tau_{\Delta_3} - 0.001714. \tag{7}$$




Figure 5. Results of the three motors' characterization ([a] first, [b] second, and [c] third motors): position (rad)–torque (Nm)–current (A) fitting (brown points and colored surfaces), relative torque (Nm)–current (A) projection (gray points), and resulting first-order polynomial fitting (blue line) with relative equation.

Furthermore, we performed a step response analysis. Figure 6 shows the results of the actuators response to a torque step $\tau_s = 1$ N, with a control loop frequency of 400 Hz (cycle time 2.5 ms). T_n indicates the time required by the actuators to reach the $n\%$ of the step. The times reported in the figure are the slowest between all the motors. The Rise Time T_r , defined as the time required by the response to rise from 10 to 90% of its final value ($T_r = T_{90} - T_{10}$), is $T_r = 0.0614$ s. Furthermore, T_s is the Settling Time and corresponds to T_{95} and results $T_s = 0.0728$ s. Considering the system of first order, we can then find the control loop bandwidth ω_H as

$$\omega_H = \frac{3}{T_s} = 41.2088 \frac{\text{rad}}{\text{s}}. \tag{8}$$

Psychophysical Assessment

By means of psychophysical experiments, we evaluate the capacity of human participants in discriminating the different force intensity rendered by the device in different direction.

Materials

The setup used for this experiments is shown in Figure 7. The $W\Delta$ is anchored to a structure that holds it and is fixed to a table. The subjects wear the $W\Delta$, while sitting in a comfortable position, and try to maintain the straight position of Figure 7 for the whole experiment duration. Furthermore, the subjects are blindfolded and provided with an external white noise by noise-canceling over-ear headphones. The pool of subjects is composed of six participants, three males, and three females, between 25 and 35 years old.




Figure 6. Step response of the $W\Delta$ with reference torque $\tau_s = 1$ N. T_n indicates the time required by the actuators to reach the $n\%$ of the step. The times reported are the slowest between all the motors. T_s is the Settling Time and corresponds to T_{95} . T_r is the Rise Time, defined as $T_r = T_{90} - T_{10}$.




Figure 7. Setup used for the $W\Delta$ psychophysical test: the subject, seated, leans its elbow on a comfort foam and wears the $W\Delta$, and witch is supported by a custom frame. The drawn reference frame orientation corresponds to the one of the base of the $W\Delta$ (see Figure 17a) in order to facilitate the read of plots reporting the assessment results in Figure 8.

Methods

The assessment method is divided in three subexperiments, each of them concerning one direction, according to the frame of Figure 7. We used the method of constant stimuli (Gescheider, 1976): in a forced-choice task, we presented participants with two stimuli of different intensity and asked them to report which of the two was perceived as stronger. After preliminary tests, the range of the stimuli has been set [2, 4] N. Then, the Compared Stimuli (CS) are 2, 2.33, 2.67, 3, 3.33, 3.67, and 4 N, while the Reference Stimulus (RS) is 3 N, with a step of 0.33 N. The corresponding torques at the wrist for these CS are in the range [0.12, 0.24] Nm in case of minimum l_h (0.06 m), and in the range [0.16, 0.32] Nm in case of maximum l_h (0.08 m, see Table 2). Twenty repetitions of each RS–CS couple are rendered to each subject, half with the RS as first stimulus and the other half with RS as second stimulus, all in a randomized order. Each stimulus is rendered as follows: a ramp of 0.5 s increases the applied force from 0 N to the desired stimulus to avoid discontinuities, then the force is kept constantly equal to the desired stimulus for 1.5 s and finally decreased back to 0 N with another ramp of 0.5 s. Between the first and second stimuli of each pair, a pause of 1 s is used. For each direction, then, the total number of trials is 140, presented in a pseudorandom order. The seed for the pseudorandom algorithm is changed for each new subexperiment.




Figure 8. Results of the force discrimination in the different directions for each subject (first 12 boxes) and for the total data pool (bottom big box). Abscissa axes are centered at the RS ($RS = 3\text{ N}$), with range corresponding to the one of the CS ($CS \in [2, 4]\text{ N}$). Grid vertical lines—plots left and right borders included—indicate the seven used CS (see Section “Methods”). On the ordinate axes, “Portion of greater” for each CS is reported, which is the portion of trials for that specific CS in which the subject declared it greater than the RS. Each box shows results for each CS (dots) and binomial fit (curve; see Section “Data Analysis”) in each direction (red for x , yellow for y , and blue for z ; see Figures 7 and 17a). The total plot also reports point of subjective equality (PSE) and just noticeable difference (JND) for each direction.

The order of the three subexperiments (three different directions of the applied forces) is shuffled with the same method. Once the desired force is given, the corresponding motor torques are calculated through (4), and the commanded currents through (7). Participants’ answers were recorded and saved for analysis.

Data Analysis

The data resulting from the experiment of each subject, as well as the total pool, are reported in Figure 8. Abscissa axes’ values correspond to the CS used in the experiment, whereas ordinate axes’ values correspond to the portion of answers in which the subject felt the CS greater than the RS. Colored dots indicate the result for each CS in the three directions. It follows that, in the ideal case, these dots should correspond to 0 if the CS is lower than RS, 0.5 if the CS corresponds to the RS, and 1 if the CS is greater than the RS.

By means of a generalized linear model, the human testers’ sensitivity is tested in all the principal directions. The fitting is performed assuming a binomial distribution. After the fitting, the PSE and the JND are calculated for each direction and subject, as well as for the whole pool of recorded data. The resulting fitting curves, PSE and JND values for each subject and for the total pool, are reported in Figure 8.

The Weber constant resulting from the characterization ($JND/RS \approx 10\%$) is quite close to the intervals found in Pang et al. (1991) and Jones (1989) (c.a. $[5 - 10]\%$), confirming that the $W\Delta$ does not alter the capacity of the user to discriminate forces. Furthermore, the average JND along the different direction is 0.338 N and exactly corresponds to the wrist *average force control resolution* calculated in Tan et al. (1994). This means that the device potentially allows the user to control a remote/virtual force through the device with the same accuracy they would have during a real physical interaction.

Teleoperation Framework

In this work, we implemented a classic position-force architecture (Hannaford and Ryu, 2002). The follower robot is moved through an impedance control, that takes as reference the user’s hand position/twist. The interaction wrench sensed at the remote side is fed back to the user through the interface. A representation of the whole scheme is shown in Figure 9. Its details are expansively explained in Appendix B, together with the study and characterization of the haptic feedback (Section “Haptic Feedback” in Appendix B). For the sake of brevity, we here report only the final and essential results. At the leader side, the wrench applied by the $W\Delta$ to users palm W_a^{pm} corresponds to

$$W_a^{pm}(t) = C_{W\Delta}(t)W_d^{pm}(t) = C_{W\Delta}(t)W_e^{ee}(t), \tag{9}$$

where W_d^{pm} is the desired wrench at the palm, corresponding to the environment wrench sensed at the follower end effector W_e^{ee} . $C_{W\Delta}$ is the map from the desired to the applied Wrench, which is the function of the device actuation and kinematics. Please refer to Section “Haptic Feedback” in Appendix B for its complete derivation. The follower, instead, is controlled in impedance, as said before. The total commanded Wrench $W_{f,tot}^{ee}$ expressed in the end-effector frame Ψ_{ee} is

$$W_{f,tot}^{ee}(t) = W_{f,cmd}^{ee}(t) + W_{f,cmp}^{ee}(t) = \Sigma_c(T_{pm}^{pm,0}(t), T_{ee}^{ee,0}(t)) + \hat{\Sigma}_f(T_{ee}^{ee,0}) \tag{10}$$




Figure 9. Teleoperation control loop block representation. Apices indicating the reference frame in which the entities are expressed and not reported for the sake of readability.

which is the sum of the impedance control Wrench $W_{f,cmd}^{ee}$ and the equivalent gravity compensation Wrench $W_{f,cmp}^{ee}$. T_i^{kj} indicates the twist of the frame Ψ_i with respect to Ψ_j expressed in the frame Ψ_k .

At the left side, the user/leader-side scheme is depicted. Z_* are impedances in the JS. Block $C_{W\Delta}$ corresponds to (9), while J_{wr} is the wrist Jacobian, which is associated to H_{pm}^{wr} of (21) and is used to map the torques at the wrist joint τ_{wr} in the wrench W_{wr} in the frame $\Psi_{pm} \equiv \Psi_{p\Delta}$.

At the right side, the remote/follower-side block representation is shown. Σ_* are impedances in the Cartesian. The sum of contributions from blocks Σ_c (impedance controller) and $\hat{\Sigma}_f$ (dynamic compensation) correspond to the total commanded wrench $W_{f,tot}$ of (10). This and the environment interaction force W_e influence the follower dynamics Σ_f , here indicated with its inverse due to causality consistency.

Experimental Validation

This section reports some simple experiments to validate the device used in the teleoperation framework described in the previous section. First, the experimental setup is introduced, both for the leader side and the follower side (Section “Validation Setup”). Section “Examples of Use” reports the examples of usage of the $W\Delta$ in free-space position and simple interaction and analyzes the results.

Validation Setup

The leader–follower framework used for the experimental campaign is shown in Figure 10. The follower is a Panda robotic arm by Franka Emika,⁸ the same used for the $W\Delta$ characterization (Section “Characterization”). The used follower end effector is the anthropomorphic Pisa/IIT SoftHand (Catalano et al., 2014).

The follower end-effector transformation matrix H_{ee}^0 and twist $T_{ee}^{e,0}$ are provided by the software library interface of the robot (Franka Control Interface), as well as its Jacobian J_s^{ee} and external interaction wrench W_{ext}^{ee} , used to provide the feedback to the user and to evaluate the subjects performances, as it will be explained hereafter. Since the measured force is quite noisy, a dead zone of ± 2.5 N is applied along all the three directions, followed by a classic first order discrete exponential filter with $\alpha_f = 0.8$, resulting in f_{ext}^{ee} . The desired torques at the Delta motors $\tau_{\Delta,d}$ are then calculated through (4). The stiffness values used




Figure 10. Setup used for the $W\Delta$ experimental validation.

⁸ <https://www.franka.de>.

in the follower impedance control (39) are set to 300 N/m for the linear components and 15 Nm/rad for the angular components. The used motors are not powerful enough to cover all the range of force that can be sensed during the robot–environment interactions, and indeed the device prototype is designed to assure a maximum force of 5 N along any direction for each point of the interested workspace. As shown in (4), the desired forces are directly related to the motor torques through the device Jacobian. If a motor saturates due to its current limit, the torque applied by that motor is saturated as well, and consequently the rendered force could misalign with respect to the desired one. To prevent this undesired behavior, a saturation is done dynamically on the vector of the desired currents $I_{\Delta,c}$. Given a current limit $I_{lim} = 1.5$ A, we perform a saturation as

$$\bar{I}_{\Delta,c} = I_{\Delta,c} r_{sat} = I_{\Delta,c} \frac{I_{lim}}{\max(\|I_{\Delta,c}\|_1, I_{lim})}, \tag{11}$$

where $\|\cdot\|_1$ is the one norm of the vector, that is, the absolute value of the largest component; $\bar{I}_{\Delta,c}$ is the vector of the direction consistent saturated currents; r_{sat} is the scaling factor due to saturation. The actuated torques $\bar{\tau}_{\Delta,c}$ are the one corresponding to the saturated current, calculated then through the inverse of the relations of (7). With this saturation, the force actually rendered to the user \bar{f}_a^{pm} (see (35)) maintains the same direction of f_{ext}^{ee} but is scaled of factor r_{sat} due to the saturation. Indeed,

$$\bar{f}_a^{pm} = J_{\Delta}^{pm\top} \bar{\tau}_{\Delta,c} = J_{\Delta}^{pm\top} \tau_{\Delta,c} r_{sat} = f_{ext}^{ee} r_{sat}. \tag{12}$$

The whole framework is implemented in ROS,⁹ with the Franka node running at 1 kHz, and the WA node at 400 Hz.

Examples of Use

Here, two simple usage examples of the device are reported, to show its basic functionalities. The first one is a free-space motion example, in which the user executes random movements in the environment without interacting with any object. Figure 11 reports the leader and follower end-effectors’ trajectories, the follower position error, the force applied by its impedance controller, and the environmental forces sensed and rendered to the user after the saturation (indicated as f and \bar{f} , respectively, and so avoiding the apex and the subscript used before for the sake of simplicity).

In the second example, the user executes random movements in the environment and eventually hit an object. Figure 12 depicts the data of three consecutive contacts as well as a frame extrapolated from the video of an instant in which the contact occurs. The time intervals relative to the contacts are highlighted with light blue strips. In the picture at the bottom of Figure 12, the y - and z -axes of Ψ_{pm} are reported, oriented in the same way of Ψ_{pm} , to help the reader understand the direction of the forces in the scene.

Figure 12 shows the results of a contact example. In the last plot, it is possible to see that the force rendered to the user (continuous lines) are lower than the ones sensed at the follower side (dashed lines), due to the saturation described before (Section “Validation Setup”). Note that the three components of the force are scaled by the same factor, so to maintain the same direction of the real remote interaction force. Forces are mainly distributed along y , because it is expressed in the palm frame Ψ_{pm} (see Figure 17).

The free-space motion example results (Figure 11) show the general trajectory tracking performance of the built framework. As it is possible to see, the position error (the second plot of Figure 11) is always relative low, almost null in steady state and above 0.1 m only during fast movements. The force commanded by the impedance controller (third row) follows the same pattern. In the bottom plot, forces sensed at the follower side are not null, due to imperfect force estimation in the robot arm.

⁹<https://www.ros.org>.




Figure 11. Free-space motion: trajectories and forces. For the sake of simplicity, only the linear components are reported. First row: desired follower positions (dashed) and actual ones (continuous). Second row: follower position errors. Third row: linear forces commanded by the follower impedance controller, expressed in the end-effector frame Ψ_{ee} . Fourth row: environment forces received by the W Δ (dashed) and the ones rendered to the user after the current saturation (continuous), expressed in the palm frame Ψ_{pm} .

Usability Assessment

Experiment Description

The setup used in this assessment is the same described in Section “Validation Setup.” The conducted multisubject campaign (12 subjects, 9 males, and 3 females, between 25 and 35 years old) has the objective to preliminary assess the W Δ usability and the effectiveness of the provided haptic feedback. The subjects involved in the multisubject experimental campaign are provided by two different leader interfaces: a “nonhaptic” interface, and the wearable haptic one composed by the W Δ (depicted in Figure 2).

The “nonhaptic” interface is composed by three Myo armbands, worn at the upper arm, forearm, and hand, respectively, as used in Laghi et al. (2018), resulting in a user interface with a total weight of 0.279 Kg (93 g per armband). Knowing a priori the arm link lengths, the IMUs of the three Myos are used to retrieve the relative rotation matrices and reconstruct the hand pose (see (21)).

The task is a contact and recognition action, similar to the one described in Section “Examples of Use.” A plane obstacle is put parallel to the floor ($x-y$ plane). Each subject is asked to teleoperate the follower robot with a constant speed until they recognize the contact with the obstacle. As soon as the subject is sure that the contact happened, they are asked to move away from the obstacle immediately. This task is repeated three times for each leader interface by each subject. Subjects are provided with an external white noise through noise-canceling over-ear headphones, but their vision is not occluded. Subjects are asked to




Figure 12. Contact with a surface experiment: trajectories and forces. The plot and data distribution follow the same convention of Figure 11. The light blue strips indicate the time intervals in which the contacts occur. In this case, the effect of motor current saturation on the exerted force is clearly appreciable in the fourth plot, where the force rendered to the user after the current saturation (continuous line) is way lower than the one received by the remote side (dashed line).

complete the experiment with both the “nonhaptic” and the leader interfaces, after a training session of 10 min circa for each of them. All the participants have to repeat a task three times for each leader interface–experiment combination. The order in which the two leader interfaces are proposed to each subject is randomized. Furthermore, the subjects do not receive any feedback during the experiment, to avoid any influence in their behaviors and performances.

A subjective evaluation of the device is performed through two questionnaires filled out by all subjects. The first one is a 7-point Likert scale custom questionnaire with six statements, reported in Table 4. The distribution of the answers relative to each statement of the custom questionnaire is reported in the bar plot of Figure 13 in the form of discrete violin plots, together with their mean, median, and mode values. The second questionnaire is the classic System Usability Scale (SUS; Table 5) for both interfaces, used to assess the their usability. For the definition and the calculation of the SUS score, please refer to Brooke (1996). Figure 14 shows the distribution of the answers of each SUS statement for both interfaces.

To further investigate the results of the task executed without any haptic feedback and with the feedback provided by the $W\Delta$, the maximum value of the remote interaction force along the contact direction y of the end-effector frame Ψ_{ee} ($f_{e,max}^{ee,y}$) is analyzed. The bars of Figure 15 show the average values of $f_{e,max}^{ee,y}$ for each subject, $\bar{f}_{e,max}^{ee,y}$, with both configurations (with and without $W\Delta$ feedback), as well as the average contact force along all of them (last bar group). Figure 16 shows the ratio between the recorded maximum average force with the $W\Delta$ and the nonhaptic interface, from Figure 15, a measure that better takes into account the sensitivity of the single subject.

Table 4. The custom 7-point Likert scale questionnaire for the comparison of the two leader interfaces

Statement
C.Q1 It was easier to move the robotic counterpart with the WΔ than the nonhaptic interface.
C.Q2 It was less physically tiresome to use the WΔ than the nonhaptic interface.
C.Q3 It was less psychologically tiresome to use the WΔ than the nonhaptic interface.
C.Q4 It was easier to recognize a contact with the WΔ than the nonhaptic interface.
C.Q5 It was easier to accomplish the contact and recognition task with the WΔ than the nonhaptic interface.




Figure 13. Discrete violin plots showing the distribution of the subjects' answers to the custom questionnaire reported in Table 4. Each violin is completed with its mean (black line), median (yellow line), and mode values (dotted light-blue line). Results above 4 (green area) are in favor of the WΔ, and the ones below 4 (red area) are in favor of the nonhaptic interface.

Results Analysis and Considerations

The custom questionnaire results (Table 4 and Figure 13) highlight a dichotomy: on one side, the usage of WΔ results inconvenient with respect to the nonhaptic interface in terms of pure motion control of the robotic counterpart (C.Q1) and of physical and cognitive load (C.Q2 and C.Q3). On the other, the users appreciate its usage for what concern the discrimination and regulation of the remote interaction (C.Q4 and C.Q5).

The negative result of C.Q1 can be easily justified: when the task consists simply in moving the robotic arm without concerns about interaction with the environment, it is easier to do it with a lighter interface. In this sense, the subjects prefer the nonhaptic interface, which is simpler, lighter, and less cumbersome than the WΔ. The results of C.Q2 and C.Q3 are an obvious consequence of this last observation, since more complex and heavier the interface is the higher the demanded physical and cognitive loads are.

Final scores of the SUS questionnaire also highlight an evident issue in the usage of the proposed haptic display: indeed, while the one relative to the nonhaptic setup scored a total of 73.75/100, the WΔ is 41.87/100, considerably higher. Even if the SUS author, always in Brooke (1996), says that “scores for individual items (of the scale) are not meaningful on their own,” we can use such individual scores as a tentative qualitative confirmation an insight of the custom questionnaire results. And, indeed, S.Q2 tells us that subjects felt the system complex, and S.Q3 that they found the system hard to use. Nonetheless, the subjects found the WΔ not too much cumbersome (S.Q8) and quite intuitive to use (S.Q10).

General comments collected from the subjects after experiments disclose two main reasons related to comfort issues in the WΔ usage: the first is its weight, which impedes the device usage for a long single session; the second, and most important, is the anchor to the human body: despite the presence of the elbow brace that strengthens the anchor, the Delta base is still able to tilt with respect to the forearm, since it is attached to the skin that is a soft tissue. Since the position sent to the follower depends on the recorder position of the WΔ, these device–forearm relative movements create inconsistency between the actual

Table 5. The System Usability Scale questionnaire (Brooke, 1996)

Statement	
S.Q1	I think that I would like to use this system frequently.
S.Q2	I found the system unnecessarily complex.
S.Q3	I thought the system was easy to use.
S.Q4	I think that I would need the support of technical person to be able to use this system.
S.Q5	I found the various functions in this system were well integrated.
S.Q6	I thought there was too much inconsistency in this system.
S.Q7	I would imagine that most people would learn to use this system very quickly.
S.Q8	I found the system very cumbersome to use.
S.Q9	I felt very confident using the system.
S.Q10	I needed to learn a lot of things before I could get going with this system.




Figure 14. Discrete violin plots showing the distribution of the subjects' answers to the SUS questionnaire (Table 5). The answers are grouped for each statement, with the ones relative the nonhaptic interface usage on the left (in blue) and the ones relative to the $W\Delta$ on the right (in orange). The mean (black line), median (yellow line), and mode (dotted light-blue line) values are reported for each subgroup. Results lying in the green areas are in favor of the relative interface. At the top horizontal axis, the p-values resulting from the Kruskal–Wallis test on the relative sets of answers (nonhaptic vs. $W\Delta$) are reported (green if $p \leq 0.05$, red otherwise).




Figure 15. Contact recognition multisubject experiment: average maximum force along the contact direction (y-axis) for each subject and the total one (last bars; x-axis), with (orange) and without (blue) the $W\Delta$ haptic interface. At the top horizontal axis, the p-values resulting from the Kruskal–Wallis test on the relative sets of $f_{e,max}^{e,y}$ (nonhaptic vs. $W\Delta$) are reported (green if $p \leq 0.05$, red otherwise).




Figure 16. Ratio of the average maximum forces $\bar{f}_{e,max}^{ee,y}$ between the $W\Delta$ and the nonhaptic interface cases (namely $W\Delta$ /nonhaptic from Figure 15). Results below 1 (green area) are in favor of the $W\Delta$, and the ones above 1 (red area) are in favor of nonhaptic interface.

position of the user hand and the one sent to the follower robot, as well as between the direction of the force sensed at the follower end effector and the one rendered to the user.

As said before, if on the comfort side, the $W\Delta$ is not preferred to the nonhaptic interface, it has been evaluated positively for the recognition and regulation of the remote interaction (C.Q4 and C.Q5 of Figure 13). To have an insight of these results, we had a look on the difference in the maximum interaction forces with the environment using the interfaces along the contact direction y of the end-effector frame Ψ_{ee} $\bar{f}_{e,max}^{ee,y}$ (which is the direction of the interaction) shown in Figures 15 and 16. We can see that almost all the subjects performed better with the $W\Delta$ than with the nonhaptic interface. Indeed, all the average force ratio of each subject depicted in Figure 16 (except for Subject 10) are lower than 1 (lying in the green area of the plot of Figure 16), meaning that the device helps them in recognizing faster and more effectively the contact with the surface. Indeed, in average, the $W\Delta$ allows to decrease the contact force from 18.3 N to 11.6 N (“TOT” group bar of Figure 15), with a drop of almost 35%. In some cases, the drop even exceeded the 50% (Subjects 4, 5, and 8). The p -values resulting from Kruskal–Wallis test performed for each subject and the total case (top horizontal axis of Figure 15) confirm the statistical difference of the results obtained from the two user configurations, except for the three subjects.

Discussion, Limitations, and Future Developments

From the experiment results, it is clear that the main drawback of the $W\Delta$ is its weight. Indeed, subjects prefer a lighter and less bulky interface, as it unloads physical and cognitive burden, a fortiori if no particular remote interaction is foreseen. Observations and comments from the subjects stated also that the device anchor on the arm is not firm enough to assure a comfortable usage, and it can confuse the user due to its relative moments with respect to the limb. Despite this, subjects did not complain any drawback effects regarding the reaction forces at the forearm anchors. The reason of this could be that the forearm is likely less sensitive than the wrist. Another possible reason is that since the anchor is not firm enough, relative movements have a more prominent effect than reaction forces. On this aspect, it will be important to perform a fine and complete characterization of the forces felt by the user in correspondence of the elbow brace and device base, as well as the pressures required to keep the device attached to the arm. Since the $W\Delta$ is the first of its type, these results are essential to understand the main challenges in the design of a usable solution, and hint the direction and guidelines for future works on the mechanical side:

- i. The reduction and redistribution of weight. This could include, for example, the usage of hollow rods for the legs and the shrink of ABS parts, under the conditions that its robustness and solidity is not decreased. Another more sophisticated and expansive option could be the substitution of ABS with carbon fiber (as done in Schiele and Hirzinger (2011)). Probably, it could be also possible to remotize the electronic boards (gray boxes in Figure 3a), foreseeing a belt to mount them on, then

Table 6. Maximum torque/force comparison between different small operational space and wrist KHDs: Omega Delta by Phantom Force (OD), Phantom Touch by 3D Systems (PT), Open Wrist (OW), Wres (WR), Rice Wrist (RW), and the presented Wearable Delta (WΔ) with peak and max used values (from the second and last columns of Table 3, respectively). (c) and (p) indicate if the values can be maintained continuously or for a limited amount of time (peak)

	OD	PT	OW	WR	RW	WΔ (data-sheet peak)	WΔ (max. used)
Max torque (Nm)			3.6(p)	6.52(p)	5.3(p)	2.751 ^a (p)	0.35^a (c)
Max force (N)	20(c)	3.3(c)				39.5 ^a (p)	5.5^a (c)

^aThis is the maximum force assured in all the interested workspace, as specified in Section “Device Dimensioning and Anthropomorphic Matching.” This means that, depending on the device Jacobian (and thus on its joint position), it is also possible to reach greater values.

removing their weight from the total load carried by the user arm. It could be worth also to create different device sizes, so to decrease the relative effort especially for minute users. Naturally, a big portion of the weight is due to actuators. Therefore, the best solution about this would be to remotize them as well, foreseeing a cable transmission (as done, e.g., always in Schiele and Hirzinger (2011)). Nonetheless, this solution is hard to implement and the saved weight would be at the expense of transmission efficiency and, consequently, of the device transparency. On the actuator side, then, it is worth considering to find lighter solution, taking into account the possibly loss of power or the increment of costs.

- ii. The creation of a more rigid and firm arm anchor that overcome the present issues without affecting the motion of the arm, also considering the outcomes of the characterization of needed pressures mentioned above. Regarding pressure, the results reported in Schiele (2008) are a perfect guideline, as they demonstrated that the optimum interface pressure between exoskeleton and human arm is 20 mmHg. Moreover, a possible improvement could be the design of a wider anchor, as it would reduce the feeling of applied pressure, as stated in Jarrassé and Morel (2011). For the arm anchor rigidity, improvements could be achieved using different commercial elbow braces, as well as consider ad hoc materials (rubbers, tissues, etc.) with great grip on the skin. Furthermore, the relative movements between the Delta base and the forearm could be reduced through an anchor redesign that follows the guidelines reported in Jarrassé and Morel (2011), which suggests to substitute single contact points to a couple of them, in order to produce couple of forces instead of moments, with a potential reduction of relative movements.

These improvements, if successfully achieved, will not only augment the users comfort, but also allow the usage of the WΔ for more complicated and structured tasks.

Apart from the comfort aspect, subjects appreciate the WΔ usage during remote interaction, and evaluate it effective in rendering contact forces. Indeed, an analysis of the interaction revealed that the usage of the device helps decreasing substantially the contact forces, objectively confirming its effectiveness.

Some observations are needed regarding force capabilities. Indeed, when compared with other wrist KHDs, the WΔ falls in performance. Table 6 compares the peak force and torque values of the WΔ with other small workspace and wrist KHDs. Both peak values from data sheet and the used ones (corresponding to the second and the last columns of Table 3) are reported. From this comparison, it is clear that the WΔ lacks in power capabilities, even when the peak values from data sheet are taken into account. Consequently, first, it will be necessary to exploit the whole actuator capabilities, implementing a dynamic current saturation control that allows the reach of peak currents for a small amount. This will allow the WΔ to approach other device performances. If this is not enough, the only other option is to adopt more powerful actuators, with the risk of increasing the device weight and go against the requirements for usability and comfort listed above. Furthermore, it is necessary to fine-tune the torque control in order to augment its bandwidth. Once the device and its control are optimized, we plan to evaluate the dynamic range of impedances it can render and measure its Z-width (Colgate and Brown (1994)) to fully qualify its range of haptic capabilities.

Apart this missing information, on the haptic side, it is interesting to observe that the magnitude of the total average maximum force of [Figure 16](#) is quite close to the ones recorded in [Laghi et al. \(2019\)](#) for the same contact recognition task but with the usage of a grounded interface, even if with a different teleoperation framework. This last observation raises a question on the adoption and the effectiveness of grounded interfaces with respect to the wearable ones. Indeed, the KHD interface used in [Laghi et al. \(2019\)](#) differs from the W Δ not only because it is grounded, but also for its power, substantially greater than the W Δ (they were KUKA LWR4+ and Franka Emika's Panda arms). Therefore, a comparative study between grounded and wearable interfaces, including not only the W Δ , but possibly all kinds of devices listed in Section "Introduction" and [Table 1](#), could give us an insight on the real need and effectiveness of such different devices depending, for example, on the task complexity, environmental condition (quality and delay of the communication), needed telepresence level, and so forth. The result of such study could be used as guideline for developers, and help them in choosing the better option depending on the specific needs and application.

Conclusion

At the beginning of this paper, a brief review of the existing KHDs and their characterization underlined the missing of some solutions (for full-arm and wrist) of the WOS type.

In order to partially fill the lack, the introduction of a new WOS wrist interface, called "W Δ ," followed. The W Δ is anchored to the base of the forearm and renders forces directly on the hand back. To avoid movement obstruction of the hand with respect to the forearm, the W Δ is provided with six DoFs, three given by an active Delta robot structure and three by a passive gimbal, linked in series with the Delta. A kinematic analysis of this new device has been performed, and it has been shown how it is possible to integrate and use it in the kinematic of the user's arm, useful for the hand trajectory tracking and its application in teleoperation scenarios. Then, the mechanical design and realization of the W Δ has been explained. In this phase, particular attention has been put in dimensioning a solution able to fit a large range of arm sizes. The result is a device that can be fit by any person with a height between [1.6, 2] m, involving just their forearm and hand and then leaving a large freedom of motion to the limb. A device characterization and a psychophysics test followed. The latter confirmed that the usage of the W Δ does not alter the ability of the user to discriminate forces, as well as the human average force control resolution for the wrist. A position-force control framework built for bilateral teleoperation using the W Δ has been described, and practical experimental examples showed its usage. A multisubject campaign allowed to show general function of the device and assess its usability followed. The results of these experiments have been finally used to understand the W Δ strengths and limitations, and to access the necessary improvements, also in comparison with the performances of other existing solutions. Furthermore, last observations revealed the need of broad and complete comparison between all the existing KHDs, not only limited to wrist ones, in order to understand which one is preferable depending on the specific application and performance requirements, and to help future makers in choosing the right kind of device to develop.

Acknowledgments. The authors would like to thank Michele Maimeri for the support and the effort he put in the software design.

Authorship Contributions. Conceptualization, M.L., M.G.C., and A.B.; Design, M.L. and M.G.C.; Methodology, M.L., M.G.C., G.G., and A.B.; Data curation, M.L. and G.G.; Data visualization, M.L. and G.G.; Writing original draft, M.L., M.G.C., G.G., and A.B. All authors approved the final submitted draft.

Supplementary Materials. To view supplementary material for this article, please visit <http://dx.doi.org/10.1017/wtc.2021.4>.

A video file showing the functioning of the W Δ and the various tasks analyzed in the Experimental Validation is provided.

Funding Statement. This research received no specific grant from any funding agency, commercial, or not-for-profit sectors.

Competing Interests. The authors declare no competing interests exist.

Data Availability Statement. Please contact authors for the data recorded and used in Section “Experimental Validation.”

Ethical Standards. The authors assert that all procedures contributing to this work comply with the ethical standards of the relevant national and institutional committees on human experimentation and with the Helsinki Declaration of 1975, as revised in 2008.

References

- Agarwal P, Fox J, Yun Y, O'Malley MK and Deshpande AD (2015) An index finger exoskeleton with series elastic actuation for rehabilitation: design, control and performance characterization. *The International Journal of Robotics Research* **34**(14), 1747–1772.
- Andrikopoulos G, Nikolakopoulos G and Manesis S (2015) Motion control of a novel robotic wrist exoskeleton via pneumatic muscle actuators. In *2015 IEEE 20th Conference on Emerging Technologies & Factory Automation (ETFA)*. IEEE, pp. 1–8.
- Barsotti M, Stroppa F, Mastronicola N, Marcheschi S and Frisoli A (2018). Teleoperated bilateral-arm rehabilitation with alex rehab station. In *International Conference on NeuroRehabilitation*. New Your, USA: Springer, pp. 185–189.
- Basteris A, Contu S, Plunkett TK, Kuah CW, Konczak IJ, Chua KS and Masia L (2018) Robot-aided bimanual assessment of wrist proprioception in people with acute stroke. In *2018 7th IEEE International Conference on Biomedical Robotics and Biomechatronics (Biorob)*. IEEE, pp. 473–478.
- Brooke J (1996) Sus-a quick and dirty usability scale. In Jordan PW, Thomas B, Weerdmeester BA and McClelland IL (eds), *Usability Evaluation in Industry*. London: Taylor & Francis, pp. 189–194.
- Buongiorno D, Sotgiu E, Leonardi D, Marcheschi S, Solazzi M and Frisoli A (2018) WRES: a novel 3 DoF WRist ExoSkeleton with tendon-driven differential transmission for neuro-rehabilitation and teleoperation. *IEEE Robotics and Automation Letters* **3**(3), 2152–2159.
- Catalano M, Grioli G, Farnioli E, Serio A, Piazza C and Bicchi A (2014) Adaptive synergies for the design and control of the Pisa/IIT SoftHand. *The International Journal of Robotics Research* **33**(5), 768–782.
- Choi I, Hawkes EW, Christensen DL, Ploch CJ and Follmer S (2016) Wolverine: a wearable haptic interface for grasping in virtual reality. In *2016 IEEE/RSJ International Conference on Intelligent Robots and Systems (IROS)*. IEEE, pp. 986–993.
- Clavel R (1988) Delta, a fast robot with parallel geometry. In Burckhardt CW (ed), *Proceedings of the 18th International Symposium on Industrial Robots*. New York: Springer-Verlag, pp. 91–100.
- Colgate JE and Brown JM (1994) Factors affecting the z-width of a haptic display. In *Proceedings of the 1994 IEEE International Conference on Robotics and Automation*. IEEE, pp. 3205–3210.
- Della Santina C, Piazza C, Gasparri GM, Bonilla M, Catalano MG, Grioli G, Garabini M and Bicchi A (2017) The quest for natural machine motion: an open platform to fast-prototyping articulated soft robots. *IEEE Robotics & Automation Magazine* **24**(1), 48–56.
- Esmaili M, Guy S, Dailey WD, Burdet E and Campolo D (2013) Subject-specific wrist model calibration and application to ergonomic design of exoskeletons. *IEEE Sensors Journal* **13**(9), 3293–3301.
- Fontana M, Dettori A, Salsedo F and Bergamasco M (2009) Mechanical design of a novel hand exoskeleton for accurate force displaying. In *2009 IEEE International Conference on Robotics and Automation (ICRA'09)*. IEEE, pp. 1704–1709.
- Frisoli A, Rocchi F, Marcheschi S, Dettori A, Salsedo F and Bergamasco M (2005) A new force-feedback arm exoskeleton for haptic interaction in virtual environments. In *2005 First Joint Eurohaptics Conference and Symposium on Haptic Interfaces for Virtual Environment and Teleoperator Systems. World Haptics Conference*. IEEE, pp. 195–201.
- Frisoli A, Simoncini F, Bergamasco M and Salsedo F (2007) Kinematic design of a two contact points haptic interface for the thumb and index fingers of the hand. *Journal of Mechanical Design* **129**(5), 520–529.
- Frisoli A, Sotgiu E, Avizzano C, Checcacci D and Bergamasco M (2004) Force-based impedance control of a haptic master system for teleoperation. *Sensor Review* **24**(1), 42–50.
- Gabardi M, Solazzi M, Leonardi D and Frisoli A (2018) Design and evaluation of a novel 5 DoF underactuated thumb-exoskeleton. *IEEE Robotics and Automation Letters* **3**(3), 2322–2329.
- Gescheider GA (1976) *Psychophysics: Method and Theory*. Hillsdale, NJ: Lawrence Erlbaum.
- Gupta A and O'Malley MK (2006) Design of a haptic arm exoskeleton for training and rehabilitation. *IEEE/ASME Transactions on Mechatronics* **11**(3), 280–289.
- Gupta A, O'Malley MK, Patoglu V and Burgar C (2008) Design, control and performance of RiceWrist: a force feedback wrist exoskeleton for rehabilitation and training. *The International Journal of Robotics Research* **27**(2), 233–251.
- Guthart GS and Salisbury JK (2000) The intuitive/sup TM/telesurgery system: overview and application. In *Proceedings 2000 ICRA. Millennium Conference. IEEE International Conference on Robotics and Automation. Symposia Proceedings (Cat. No. 00CH37065)*, Vol. 1. IEEE, pp. 618–621.
- Hannaford B and Ryu J-H (2002) Time-domain passivity control of haptic interfaces. *IEEE Transactions on Robotics and Automation* **18**(1), 1–10.
- Hope J and McDaid A (2017) Development of wearable wrist and forearm exoskeleton with shape memory alloy actuators. *Journal of Intelligent and Robotic Systems* **86**(3–4), 397–417.

- Hulin T, Hertkorn K, Kremer P, Schätzle S, Artigas J, Sagardia M, Zacharias F and Preusche C** (2011). The DLR bimanual haptic device with optimized workspace. In *2011 IEEE International Conference on Robotics and Automation (ICRA)*. IEEE, pp. 3441–3442.
- Imaida T, Yokokohji Y, Doi T, Oda M and Yoshikawa T** (2004) Ground-space bilateral teleoperation of ETS-VII robot arm by direct bilateral coupling under 7-s time delay condition. *IEEE Transactions on Robotics and Automation* **20**(3), 499–511.
- Iqbal J, Tsagarakis N and Caldwell D** (2015) Four-fingered lightweight exoskeleton robotic device accommodating different hand sizes. *Electronics Letters* **51**(12), 888–890.
- Jarrassé N and Morel G** (2011) Connecting a human limb to an exoskeleton. *IEEE Transactions on Robotics* **28**(3), 697–709.
- Jones LA** (1989) Matching forces: constant errors and differential thresholds. *Perception* **18**(5), 681–687.
- Khatib O, Yeh X, Brantner G, Soe B, Kim B, Ganguly S, Stuart H, Wang S, Cutkosky M, Edsinger A, Mullins P, Barham M, Woolstra CR, Salama KN, L'Hour M and Creuze V** (2016) Ocean one: a robotic avatar for oceanic discovery. *IEEE Robotics & Automation Magazine* **23**(4), 20–29.
- Klamt T, Rodriguez D, Baccelliere L, Chen X, Chiaradia D, Cichon T, Gabardi M, Guria P, Holmquist K, Kamedula M, Karaoguz H, Kashiri N, Laurenzi A, Lenz C, Leonardi D, Hoffman EM, Muratore L, Pavlichenko D, Porcini F, Ren Z, Schilling F, Schwarz M, Solazzi M, Felsberg M, Frisoli A, Gustmann M, Jensfelt P, Nordberg K, Roßmann J, Süß U, Tsagarakis NG and Behnke S** (2019a) Flexible disaster response of tomorrow: final presentation and evaluation of the centauro system. *IEEE Robotics & Automation Magazine* **26**(4), 59–72.
- Klamt T, Schwarz M, Lenz C, Baccelliere L, Buongiorno D, Cichon T, DiGuardo A, Droschel D, Gabardi M, Kamedula M, Kashiri N, Laurenzi A, Leonardi D, Muratore L, Pavlichenko D, Periyasamy AS, Rodriguez D, Solazzi M, Frisoli A, Gustmann M, Roßmann J, Süß U, Tsagarakis NG and Behnke S** (2019b) Remote mobile manipulation with the centauro robot: full-body telepresence and autonomous operator assistance. *Journal of Field Robotics* **37**(5), 889–919.
- Laghi M, Ajoudani A, Catalano MG and Bicchi A** (2019) Unifying bilateral teleoperation and tele-impedance for enhanced user experience. *The International Journal of Robotics Research* **39**(4), 514–539.
- Laghi M, Maimeri M, Marchand M, Leparoux C, Catalano M, Ajoudani A and Bicchi A** (2018) Shared-autonomy control for intuitive bimanual tele-manipulation. In *IEEE/RAS International Conference on Humanoids Robots*. IEEE, pp. 1–9.
- Lambelet C, Lyu M, Woolley D, Gassert R and Wenderoth N** (2017) The eWrist—a wearable wrist exoskeleton with sEMG-based force control for stroke rehabilitation. In *2017 International Conference on Rehabilitation Robotics (ICORR)*. IEEE, pp. 726–733.
- Lee J, Song B-W and Yang W** (2018) Design of exoskeleton-type wrist human–machine interface based on over-actuated coaxial spherical parallel mechanism. *Advances in Mechanical Engineering* **10**(2). <https://doi.org/10.11818/1687814017753896>
- Lee K-M and Shah DK** (1988) Kinematic analysis of a three-degrees-of-freedom in-parallel actuated manipulator. *IEEE Journal on Robotics and Automation* **4**(3), 354–360.
- Leonardi D, Barsotti M, Loconsole C, Solazzi M, Troncosi M, Mazzotti C, Castelli VP, Procopio C, Lamola G, Chisari C, Bergamasco M and Frisoli A** (2015) An EMG-controlled robotic hand exoskeleton for bilateral rehabilitation. *IEEE Transactions on Haptics* **8**(2), 140–151.
- Letier P, Avraam M, Veillerette S, Horodincu M, De Bartolomei M, Schiele A and Preumont A** (2008) SAM: a 7-DoF portable arm exoskeleton with local joint control. In *2008 IEEE/RSJ International Conference on Intelligent Robots and Systems (IROS 2008)*. IEEE, pp. 3501–3506.
- Mallwitz M, Will N, Teiwes J and Kirchner EA** (2015) The CAPIO active upper body exoskeleton and its application for teleoperation. In *Proceedings of the 13th Symposium on Advanced Space Technologies in Robotics and Automation. ESA/Estec Symposium on Advanced Space Technologies in Robotics and Automation (ASTRA-2015)*. ESA.
- Martinez JA, Ng P, Lu S, Campagna MS and Celik O** (2013) Design of wrist gimbal: a forearm and wrist exoskeleton for stroke rehabilitation. In *2013 IEEE 13th International Conference on Rehabilitation Robotics (ICORR)*. IEEE, pp. 1–6.
- Oblak J, Cikajlo I and Matjacic Z** (2010) Universal haptic drive: a robot for arm and wrist rehabilitation. *IEEE Transactions on Neural Systems and Rehabilitation Engineering* **18**(3), 293–302.
- Okamura AM, Richard C and Cutkosky MR** (2002) Feeling is believing: using a force-feedback joystick to teach dynamic systems. *Journal of Engineering Education* **91**(3), 345–349.
- Pacchierotti C, Sinclair S, Solazzi M, Frisoli A, Hayward V and Prattichizzo D** (2017) Wearable haptic systems for the fingertip and the hand: taxonomy, review, and perspectives. *IEEE Transactions on Haptics* **10**(4), 580–600.
- Pang X-D, Tan HZ and Durlach NI** (1991) Manual discrimination of force using active finger motion. *Perception & Psychophysics* **49**(6), 531–540.
- Part S** (1985) Impedance control: an approach to manipulation. *Journal of Dynamic Systems, Measurement, and Control* **107**(1), 1–7.
- Pehlivan AU, Lee S and O'Malley MK** (2012) Mechanical design of RiceWrist-S: a forearm–wrist exoskeleton for stroke and spinal cord injury rehabilitation. In *2012 4th IEEE RAS & EMBS International Conference on Biomedical Robotics and Biomechatronics (BioRob)*. IEEE, pp. 1573–1578.
- Pehlivan AU, Sergi F and O'Malley MK** (2014) A subject-adaptive controller for wrist robotic rehabilitation. *IEEE/ASME Transactions on Mechatronics* **20**(3), 1338–1350.
- Pezent E, Fani S, Clark J, Bianchi M and O'Malley MK** (2019) Spatially separating haptic guidance from task dynamics through wearable devices. *IEEE Transactions on Haptics* **12**(4), 581–593.

Pezent E, Rose CG, Deshpande AD and O'Malley MK (2017) Design and characterization of the openwrist: a robotic wrist exoskeleton for coordinated hand-wrist rehabilitation. In *2017 International Conference on Rehabilitation Robotics (ICORR)*. IEEE, pp. 720–725.

Rosen J and Ferguson PW (2020) *Wearable Robotics: Systems and Applications*. Academic Press.

Sarakoglou I, Brygo A, Mazzanti D, Hernandez NG, Caldwell DG and Tsagarakis NG (2016) Hexotrac: a highly under-actuated hand exoskeleton for finger tracking and force feedback. In *2016 IEEE/RSJ International Conference on Intelligent Robots and Systems (IROS)*. IEEE, pp. 1033–1040.

Schiele A (2008) Fundamentals of ergonomic exoskeleton robots. Ph.D. Thesis, Delft University of Technology.

Schiele A and Hirzinger G (2011) A new generation of ergonomic exoskeletons—the high-performance x-arm-2 for space robotics telepresence. In *2011 IEEE/RSJ International Conference on Intelligent Robots and Systems (IROS)*. IEEE, pp. 2158–2165.

Schiele A and Van Der Helm FC (2006) Kinematic design to improve ergonomics in human machine interaction. *IEEE Transactions on Neural Systems and Rehabilitation Engineering* **14**(4), 456–469.

Sergi F, Accoto D, Tagliamonte NL, Carpino G, Galzerano S and Guglielmelli E (2012). Kinematic synthesis, optimization and analysis of a non-anthropomorphic 2-DoFs wearable orthosis for gait assistance. In *2012 IEEE/RSJ International Conference on Intelligent Robots and Systems*. IEEE, pp. 4303–4308.

Tan HZ, Srinivasan MA, Eberman B and Cheng B (1994) Human factors for the design of force-reflecting haptic interfaces. *Dynamic Systems and Control* **55**(1), 353–359.

Williams RL (2016) The delta parallel robot: kinematics solutions. *Internal Publication*.

Winter DA (2009) *Biomechanics and Motor Control of Human Movement*. John Wiley & Sons.

Yip MC, Tavakoli M and Howe RD (2011) Performance analysis of a haptic telemanipulation task under time delay. *Advanced Robotics* **25**(5), 651–673.

A. Kinematics

This appendix analyzes the kinematics of device and the user’s arm. First, the device direct kinematics is presented in Section “WΔ kinematics” in Appendix A. Then, the user’s limb pose, which is needed for the teleoperation framework presented in Section “Teleoperation Framework,” is derived in Section “Human Arm Kinematics” in Appendix A. Since the device will be worn during teleoperation, it can be used as a sensor to measure the pose of the arm it embraces, so to avoid additional sensors. The integration of the device kinematics in the arm’s (Section “Integrating the WΔ in the Human Arm Kinematics” in Appendix A) will conclude this section.




Figure 17. Representation of the main lengths and frames of (a) the device and (b) the human arm. Ψ_* indicate the position of reference frames used in the text to define the kinematics of the WΔ and the human arm.

A.1. WA Kinematics

The $W\Delta$ has six DoFs, resulting from the serial assembly of two three-DoF structures, the Delta parallel chain (three linear DoFs) and the gimbal serial one (three angular DoFs). The homogeneous transformation $H_{pm}^{b_{W\Delta}}$ between the palm frame Ψ_{pm} and the device base frame $\Psi_{b_{\Delta}}$ is

$$H_{pm}^{b_{W\Delta}}(q_{W\Delta}) = H_{p_{\Delta}}^{b_{\Delta}}(\theta_{\Delta})H_{pm}^{p_{\Delta}}(\theta_g) = \begin{bmatrix} R_{pm}^{b_{\Delta}}(\theta_g) & P_{p_{\Delta}}^{b_{\Delta}}(\theta_{\Delta}) \\ \mathbf{0}_{(1 \times 3)} & 1 \end{bmatrix}, \tag{13}$$

where $q_{W\Delta} = [\theta_{\Delta}^T \ \theta_g^T]^T$, with θ_{Δ} being the vector of the Delta motors' position and θ_g the rotation angles vector of the gimbal; $H_{p_{\Delta}}^{b_{\Delta}}$ is the homogeneous transformation matrix between the Delta moving plate $\Psi_{p_{\Delta}}$ and the base frame $\Psi_{b_{\Delta}}$; $H_{pm}^{p_{\Delta}}$ is between the palm frame Ψ_{pm} and the moving plate frame $\Psi_{p_{\Delta}}$. As highlighted, $P_{p_{\Delta}}^{b_{\Delta}}$ depends only on the Delta motor position vector θ_{Δ} , and indeed is the result of the Delta forward kinematics. Being a parallel structure, its forward kinematics cannot be defined solely by the chain rule of homogeneous transformation. Instead, it is given by imposing position constrains between the three legs. A possible solution is that proposed in Williams (2016), which is adopted in this work. On the other side, $R_{pm}^{b_{\Delta}}$ depends only on the rotation angles of the gimbal axis θ_g (see Figure 17a):

$$R_{pm}^{b_{\Delta}}(\theta_g) = R_{x, -\frac{\pi}{2}} R_{z, \frac{\pi}{2}} R_{z, \theta_{g1}} R_{x, \theta_{g2}} R_{y, \theta_{g3}}, \tag{14}$$

where, for example, $R_{\chi, \alpha}$ is a rotation around the χ -axis of α radians. Finally, it is easy and straightforward to obtain the homogeneous transformation matrix H_{pm}^{el} between the palm and the elbow, as

$$H_{pm}^{el}(q_{W\Delta}) = H_{b_{\Delta}}^{el} H_{pm}^{b_{\Delta}}(q_{W\Delta}), \tag{15}$$

with $H_{b_{\Delta}}^{el}$ being the constant homogeneous transformation between the Delta base and the elbow frames, since the Delta is rigidly anchored to the forearm.

In Williams (2016), also a solution for the calculation of the Jacobian matrix of the Delta structure is given. Following the same procedure, the Delta Jacobian $J_{\Delta}^{b_{\Delta}}$ expressed in the base frame $\Psi_{b_{\Delta}}$ that maps the motor velocities $\dot{\theta}_{\Delta}$ into the twist $T_{p_{\Delta}}^{b_{\Delta}, b_{\Delta}}$, by

$$T_{pm}^{b_{\Delta}, b_{\Delta}} = J_{\Delta}^{b_{\Delta}} \dot{\theta}_{\Delta}, \tag{16}$$

is given by

$$J_{\Delta}^{b_{\Delta}} = \begin{bmatrix} \mathbf{0}_{(3 \times 3)} \\ V_{\Delta}^{b_{\Delta}} \end{bmatrix}_{(6 \times 3)} = \begin{bmatrix} \mathbf{0}_{(3 \times 3)} \\ A^{-1} B \end{bmatrix}_{(6 \times 3)}, \tag{17}$$

with A and B 3×3 matrices resulting from the derivation of direct kinematics equations (refer again to Williams (2016) for their calculation). Note that, since the three DoFs of the Delta structure are all linear, its actuation produces only linear velocities. This is why the upper block of $J_{\Delta}^{b_{\Delta}}$ is a 3×3 null matrix.

The gimbal Jacobian $J_g^{p_{\Delta}}$, expressed in the moving plate frame $\Psi_{p_{\Delta}}$, that maps the gimbal joint velocities $\dot{\theta}_g$ into the twist $T_{pm}^{p_{\Delta}, p_{\Delta}}$, as

$$T_{pm}^{p_{\Delta}, p_{\Delta}} = J_g^{p_{\Delta}} \dot{\theta}_g, \tag{18}$$

is given by

$$J_g^{p_{\Delta}} = \begin{bmatrix} \widehat{T}_{g1}^{p_{\Delta}, p_{\Delta}} & \widehat{T}_{g2}^{p_{\Delta}, p_{\Delta}} & \widehat{T}_{g3}^{p_{\Delta}, p_{\Delta}} \end{bmatrix}_{(6 \times 3)} = \begin{bmatrix} \widehat{\omega}_{g1}^{p_{\Delta}, p_{\Delta}} & \widehat{\omega}_{g2}^{p_{\Delta}, p_{\Delta}} & \widehat{\omega}_{g3}^{p_{\Delta}, p_{\Delta}} \\ \mathbf{0}_{(3 \times 3)} \end{bmatrix}_{(6 \times 3)} = \begin{bmatrix} \Omega \\ \mathbf{0}_{(3 \times 3)} \end{bmatrix}, \tag{19}$$

where $\widehat{\omega}$ are the unit vectors indicating the rotation axis of each of the gimbal DoF. Finally, the Jacobian of the whole device $J_{W\Delta}^{b_{\Delta}}$ is given by the combination of the Delta and the gimbal Jacobians:

$$J_{W\Delta}^{b_{\Delta}} = \begin{bmatrix} J_{\Delta}^{b_{\Delta}} & J_g^{b_{\Delta}} \end{bmatrix}_{(6 \times 6)} = \begin{bmatrix} \mathbf{0} \\ A^{-1} B \end{bmatrix} Ad_{H_{p_{\Delta}}^{b_{\Delta}}} \begin{bmatrix} \Omega \\ \mathbf{0}_{(3 \times 3)} \end{bmatrix}_{(6 \times 6)}, \tag{20}$$

with $Ad_{H_{p_{\Delta}}^{b_{\Delta}}}$ being the adjoint matrix of the homogeneous transformation $H_{p_{\Delta}}^{b_{\Delta}}$.

A.2. Human Arm Kinematics

The kinematics of the human arm can be approximated by a seven-DoF serial kinematic chain with three links (as shown in Figure 17b): the upper arm with length l_u , the forearm with length l_f , and the hand with length l_h from the wrist to the palm. The chain joints are: a spherical joint for the shoulder (three DoFs, $q_{sh}(3 \times 1)$), a rotational joint for the elbow (one DoF, q_{el}), and another spherical joint for the wrist (three DoFs, $q_{wr}(3 \times 1)$). Then, the homogeneous transformation H_{pm}^{sh} from the palm to the shoulder is given by

$$H_{pm}^0(q_H) = H_{sh}^0(q_{sh})H_{el}^{sh}(q_{el})H_{wr}^{el}(q_{wr})H_{pm}^{wr}, \tag{21}$$

where $q_{H(7 \times 1)} = [q_{sh}^T \ q_{el} \ q_{wr}^T]^T$.

A.3. Integrating the WΔ in the Human Arm Kinematics

Previously, we explained how to calculate the position and orientation of the palm with respect to the elbow through the direct kinematics of $W\Delta$. Then, since $H_{pm}^{el}(q_{wr}) \equiv H_{pm}^{el}(q_{W\Delta})$, it is possible to use $W\Delta$ to calculate a portion of the human palm pose. Including (15) in (21), we obtain

$$H_{pm}^0(q_{\bar{H}}) = H_{sh}^0(q_{sh})H_{el}^{sh}(q_{el})H_{b\Delta}^{el}H_{pm}^{b\Delta}(q_{W\Delta}), \tag{22}$$

where $q_{\bar{H}(10 \times 1)} = [q_{sh}^T \ q_{el} \ q_{W\Delta}^T]^T$ indicates the vector of joints position of human shoulder, human elbow, and $W\Delta$.

The human arm Jacobian J_H^{pm} is the matrix that maps the arm joint velocities \dot{q}_H into the twist $T_{pm}^{pm,0}$ of the palm with respect to the world expressed in a palm frame, as in

$$T_{pm}^{pm,0} = J_H^{pm} \dot{q}_H. \tag{23}$$

Moreover, $T_{pm}^{pm,0}$ can also be expressed as the sum of two twists as

$$T_{pm}^{pm,0} = T_{el}^{pm,0} + T_{pm}^{pm,el} = J_{H_{el}^{pm}}^{pm} \begin{bmatrix} \dot{q}_{sh} \\ \dot{q}_{el} \end{bmatrix} + J_{H_{pm}^{el}}^{pm} \dot{q}_{wr}, \tag{24}$$

where $J_{H_{el}^{pm}}^{pm}$ and $J_{H_{pm}^{el}}^{pm}$ are the Jacobians of the shoulder–elbow and elbow–palm chains, respectively, both expressed in the palm frame. Furthermore, since $H_{pm}^{el}(q_H) = H_{b\Delta}^{el}H_{pm}^{b\Delta}(q_{W\Delta})$, we also have

$$T_{pm}^{pm,el} = J_{H_{pm}^{el}}^{pm} \dot{q}_{wr} = J_{W\Delta}^{pm} \dot{q}_{W\Delta}, \tag{25}$$

where $J_{W\Delta}^{pm} = Ad_{H_{b\Delta}^{pm}} J_{W\Delta}^{b\Delta}$ is the $W\Delta$ Jacobian expressed in the palm frame. From these considerations, it is possible to define another Jacobian $J_{\bar{H}}^{pm}$ that maps the joint velocity vector $\dot{q}_{\bar{H}}$ always into $T_{pm}^{pm,0}$, making use of the $W\Delta$ kinematics:

$$T_{pm}^{pm,0} = J_{\bar{H}}^{pm} \dot{q}_{\bar{H}} = \begin{bmatrix} J_{H_{el}^{pm}}^{pm} & J_{W\Delta}^{pm} \end{bmatrix} \begin{bmatrix} \dot{q}_{sh} \\ \dot{q}_{el} \\ \dot{q}_{W\Delta} \end{bmatrix}. \tag{26}$$

B. Details of the teleoperation framework

The bilateral teleoperation framework used in this work is a classic position-force architecture (Hannaford and Ryu, 2002). The follower robot is moved through an impedance control, that takes as reference the user’s hand position/twist. The interaction wrench sensed at the remote side is fed back to the user through the interface. Since the developed interface does not involve all the user’s arm, but only a portion of it, an adaptation of the framework is needed. In particular, the leader side has to be split in two parts, one for the arm segments not involved in the feedback and the second for the segments involved. A representation of the whole scheme is shown in Figure 9 and will be clarified hereafter. The haptic feedback study is presented in Section “Haptic Feedback” in Appendix B, whereas the follower robot control is presented next.

B.1. Haptic Feedback

Let us consider the palm frame Ψ_{pm} , whose origin coincide with the frame $\Psi_{p\Delta}$

$$H_{pm}^{p\Delta} = \begin{bmatrix} R_{pm}^{p\Delta} & 0_{(3 \times 1)} \\ 0_{(1 \times 3)} & 1 \end{bmatrix},$$

and analyze the device Jacobian $J_{W\Delta}^{pm}$, expressed in that frame. $J_{W\Delta}^{pm}$ can be written as

$$J_{W\Delta}^{pm} = [J_{\Delta}^{pm} \ J_g^{pm}] = [Ad_{H_{b\Delta}^{pm}} J_{\Delta}^{b\Delta} \ Ad_{H_{b\Delta}^{pm}} J_g^{pm}], \tag{27}$$

with

$$J_{\Delta}^{b\Delta} = \begin{bmatrix} 0 \\ A^{-1}B \end{bmatrix}, \quad J_g^{p\Delta} = \begin{bmatrix} \Omega \\ 0 \end{bmatrix}, \tag{28}$$

and $Ad_{H_{b\Delta}^{pm}}$ and $Ad_{H_{p\Delta}^{pm}}$ being the adjoint matrices associated to $H_{b\Delta}^{pm}$ and $H_{p\Delta}^{pm}$, respectively, defined as

$$Ad_{H_{b\Delta}^{pm}} = \begin{bmatrix} R_{b\Delta}^{pm} & 0 \\ \tilde{p}_{b\Delta}^{pm} R_{b\Delta}^{pm} & R_{b\Delta}^{pm} \end{bmatrix}; \quad Ad_{H_{p\Delta}^{pm}} = \begin{bmatrix} R_{p\Delta}^{pm} & 0 \\ 0 & R_{p\Delta}^{pm} \end{bmatrix}. \tag{29}$$

Using (28) and (29) in (27), $J_{W\Delta}^{pm}$ can be written as

$$J_{W\Delta}^{pm} = \begin{bmatrix} 0 & R\Omega \\ RA^{-1}B & 0 \end{bmatrix}, \tag{30}$$

with $R \triangleq R_{b\Delta}^{pm} = R_{p\Delta}^{pm}$.

Once we obtain the device Jacobian $J_{W\Delta}^{pm}$ expressed in the palm frame Ψ_{pm} , it is possible to study the forces and torques that the device is able to render at the user’s hand. Consider a desired wrench W_d^{pm} expressed in the palm frame

$$W_d^{pm} = \begin{bmatrix} f_d^{pm} \\ \tau_d^{pm} \end{bmatrix}, \tag{31}$$

where f_d^{pm} and τ_d^{pm} are the vectors of the desired forces and torques, respectively. The desired torques $\tau_{W\Delta,d}$ to be applied at the device joints so to obtain W_d^{pm} are given by

$$\tau_{W\Delta,d} = \begin{bmatrix} \tau_{\Delta,d} \\ \tau_{g,d} \end{bmatrix} = J_{W\Delta}^{pmT} \begin{bmatrix} f_d^{pm} \\ \tau_d^{pm} \end{bmatrix} = \begin{bmatrix} 0 & B^T A^{-T} R^T \\ \Omega^T R^T & 0 \end{bmatrix} \begin{bmatrix} f_d^{pm} \\ \tau_d^{pm} \end{bmatrix}, \tag{32}$$

that yields to

$$\begin{aligned} \tau_{\Delta,d} &= B^T A^{-T} R^T f_d^{pm}, \\ \tau_{g,d} &= \Omega^T R^T \tau_d^{pm}. \end{aligned} \tag{33}$$

Two observations can be made on this last result:

- i. Being the gimbal a passive structure, we always have $\tau_{g,a} = 0$, where the subscript a stands for *actual*. This means that, as long as Ω is nonsingular (and so invertible), the actual torques applied to the palm by the device τ_a^{pm} result

$$\tau_{g,a} = 0 \Rightarrow \tau_a^{pm} = R\Omega^{-T} \tau_{g,a} = 0. \tag{34}$$

- ii. As long as B is invertible, we can control the forces applied to the palm f_a^{pm} through $\tau_{\Delta,a}$. Indeed, given a desired f_d^{pm} , we chose $\tau_{\Delta,a} = \tau_{\Delta,d} = B^T A^{-T} R^T f_d^{pm}$ as in (33), and then

$$f_a^{pm} = RA^T B^{-T} \tau_{\Delta,d} = RA^T B^{-T} B^T A^{-T} R^T f_d^{pm} = f_d^{pm}. \tag{35}$$

We can then define a new Jacobian $J_{W\Delta,a}^{pm}$, that maps the desired wrench W_d^{pm} into the torques applicable by the device joints $\tau_{W\Delta,a}$

$$\tau_{W\Delta,a} = \begin{bmatrix} \tau_{\Delta,a} \\ \tau_{g,a} \end{bmatrix} = J_{W\Delta,a}^{pmT} W_d^{pm} = \begin{bmatrix} 0 & 0 \\ RA^{-1}B & 0 \end{bmatrix}^T W_d^{pm} = \begin{bmatrix} \tau_{\Delta,a} \\ \tau_{g,a} \end{bmatrix} = \begin{bmatrix} B^T A^{-T} R^T f_d^{pm} \\ 0 \end{bmatrix}. \tag{36}$$

Finally, we have the wrench W_a^{pm} applied at the palm by the device

$$W_a^{pm} = J_{W\Delta,a}^{pmT} \tau_{W\Delta,a}, \tag{37}$$

where † indicates pseudoinversion. Combining (36) and (37), we can explicit the relation between the desired and applied wrenches, W_d^{pm} and W_a^{pm} , respectively,

$$W_a^{pm} = J_{W\Delta}^{pm \dagger T} J_{W\Delta,a}^{pm T} W_d^{pm} = C_{W\Delta} W_d^{pm} \tag{38}$$

Since the device is anchored on the forearm and on the hand back, this wrench only affects the wrist joint. Indeed, a closed loop is created only between the forearm–backhand chain (whose impedance in the JS is indicated as Z_{wr}) and the device ($Z_{W\Delta}$), whose dynamics are then affected only by the torques of the user’s wrist τ_{wr} (see Figure 9). Always from Figure 9, it is possible to see that in a teleoperation scenario, the desired wrench W_d^{pm} of (36) corresponds to W_e , which is the wrench sensed by the follower at the remote side during the interaction with the environment.

Above, the weight of the device has not been considered. Naturally, also this one is perceived by the user on its whole arm, as they has to continuously compensate for it. On this aspect, please also note that no gravity compensation is foreseen in the $W\Delta$ control loop (left side of Figure 9). This choice has been made after two considerations:

- the parts which could be compensated (which are the gimbal structure, in ABS material, and the delta legs, in stainless steel) are light compared to the total weight of the device (which is concentrated in the Delta base due to motors and electronics);
- even if the weight of such components are compensated, no particular relief to the user would have been given, since such weight is inevitably loaded on the user arm (as underlined just above).

Taking also into account the considerably complexity of an active gravity compensation control for a parallel structure, it has been then decided to not implement it.

B.2. Follower Control

The remote robotic arm (follower) is controlled through a classic Cartesian impedance control (Part 1985; see Figure 9): the desired homogeneous transformation matrix resulting from (1) is sent to the remote robot controller, that computes the desired wrench $W_{f,cmd}(t)$ as

$$W_{f,cmd}^{ee}(t) = \Sigma_c \left(T_{pm}^{pm,0}(t), T_{ee}^{ee,0}(t) \right) = K_c(t) * \tilde{x}(t) + D_c(t) * \dot{\tilde{x}}(t), \tag{39}$$

where Σ_c is the Cartesian impedance controller; $T_{pm}^{pm,0}$ comes from (2); $T_{ee}^{ee,0}$ is the twist of the follower end-effector frame Ψ_{ee} with respect to its base frame expressed in Ψ_{ee} ; $\tilde{x} = T_{pm}^{pm,0} - T_{ee}^{ee,0}$ is the velocity error vector and \tilde{x} is its time integration; and K_c is the commanded follower stiffness matrix at the end effector, while D_c is the commanded damping matrix at the end effector, in this work defined imposing a critical damping factor $\zeta = 1$. The total wrench $W_{f,tot}^{ee}$ commanded to the follower is, finally, comprehensive of the dynamic compensation terms $W_{f,cmp}^{ee}$:

$$W_{f,tot}^{ee}(t) = W_{f,cmd}^{ee}(t) + W_{f,cmp}^{ee}(t) = \Sigma_c \left(T_{pm}^{pm,0}(t), T_{ee}^{ee,0}(t) \right) + \widehat{\Sigma}_f (T_{ee}^{ee,0}) \tag{40}$$

where $W_{f,cmp}^{ee}$ is the follower dynamic compensation wrench resulting from the estimated follower dynamics $\widehat{\Sigma}_f$. The wrench of (40) is the total acting on the follower, in Figure 6 indicated in its admittance form Σ^{-1} , because of the scheme causality.

Please note that the control depicted in Figure 9 is valid under the conditions $\Sigma_f - \widehat{\Sigma}_f + \Sigma_e + \Sigma_c \neq 0$ at the follower side, and $I + J_{wr}^{-T} Z_{wr} Z_{W\Delta}^{-1} J_{W\Delta}^T \neq 0$ at the leader side.